

DELHI POLLUTION CONTROL COMMITTEE

DEPARTMENT OF ENVIRONMENT, GOVT. OF DELHI

4th& 5th FLOOR, ISBT BUILDING, KASHMERE GATE, DELHI-06

visit us at : <http://dpcc.delhigovt.nic.in>

F.No. DPCC/Lab(W)/2018/P-19/1109

Dated: 24/09/2018

To,

SEE (IT) - to upload on the DPCC website

Sub: Quotation for the purchase of COD Digester.


Sir,

Sealed quotations are invited for the purchase of COD Digester (20 holes). Specifications for the instrument is enclosed with quotation letter. Quotation must be reach to this office latest by 15.09.2018 by 4.00 PM at enquiry counter, 6th Floor, ISBT Building, Kashmere Gate, Delhi.

Terms and Conditions:

1. The Agency must Produce/ enclose registration certificate of GST.
2. The Specification mentioned in the quotation are fixed and no variation in the same shall be accepted.
3. The sealed envelope containing quotation shall be super scribed in the top of the envelope, "Quotation for the purchase of (Name of Instruments).
4. The rates should quoted for each item in the category exclusive of taxes as applicable.
5. Any item supplied of inferior quality shall be replaced without any extra charge.
6. The material should be supplied on F.O.R. basis.
7. Failure to supply the items within stipulated time shall be viewed seriously & the earnest money shall be forfeited.
8. The payment shall be made after the receipt of material / Instruments.

The Competent Authority, DPCC reserve the all right to reject or accept any quotation without any reason thereof.


(Dr. Nandita Moitra)
Senior Scientist
Dr. NANDITA MOITRA
Senior Scientist
Delhi Pollution Control Committee
5th Floor, ISBT Building,
Kashmere Gate, New Delhi-110006

Procurement of instruments/equipment for Delhi Pollution Control Board Laboratories

Code no. 1.

Equipment Name: COD DIGESTOR

Quantity: 1(one)Nos.

Sl.No.	Specifications	Requirements	BIDDER'S RESPONSE: (Bidder shall indicate "Complied" or "Non Complied" as applicable in each row. Technical Literature/ published document should be submitted to prove compliance)
.0	Description	Solid state COD digestion unit complete with required accessories for digestion of environmental samples for measurement of Chemical Oxygen Demand (COD).	
.0	Heating Unit	Solid State block heater unit accommodating 20 nos. 38 mm diameter reaction vessels	
.0	Reaction vessels	38 mm diameter glass reaction vessels with 20 ml. sample size – set of 20 nos.	
.0	Heating temperature	150 °C ± 1 °C	
.0	Temperature control	PID based Digital Controller	
.0	Temperature Indicator	LCD Display indicator with date and time	
.0	Air condenser	20 Nos. air condensers suitable for fitting in glass reaction vessels	
.0	Timer	Digital timer with buzzer	
.0	Power supply	230 ±10 V 50 Hz AC	
0.0	Accessories	Two additional sets of 20 Nos. glass reaction vessels and 20 Nos. air condensers (Optional) Calibration Certificate from NABL Accredited Lab to be provided for calibration of Temperature Control.	