



DELHI POLLUTION CONTROL COMMITTEE
DEPARTMENT OF ENVIRONMENT, GOVT. OF DELHI
4th & 5th FLOOR, ISBT BUILDING, KASHMERE GATE, DELHI-06
visit us at : <http://dpcc.delhigovt.nic.in>

No. F. DPCC/1810

Dated: 08/12/2018

To,
SEE IT (Upload to DPCC website)

Sub: Quotation for the purchase of Online Ammonia Analyser.

Sir,

Sealed quotations are invited for the purchase, Commissioning, Installation operation of following Instruments and monitoring the Ammonia level for a period of one year at two locations (Palla and Wazirabad).

1. Online Ammonia Analyser (2 Units).

Specifications for the instruments should be enclosed along with quotation letter. Quotation must be reach in this office latest by 20.12.18 at 4.00 PM at enquiry counter, 4th floor, ISBT Building, Kashmere Gate, Delhi as per following Specifications:

Principal – Ion Selective Electrode Method/ UV Spectrophotometer
Range- 0.5 mg/l to 100 mg/l
Accuracy-Minimum \pm 0.5mg/l or better

Terms and Conditions:

- 1. The Agency must Produce/ enclose registration certificate of GST.**
2. The Specification mentioned in the quotation are fixed and no variation in the same shall be accepted.
3. The sealed envelope containing quotation shall be super scribed in the top of the envelope, "Quotation for the purchase of (Name of Instruments).
4. The rates should quoted for each item in the category exclusive of taxes as applicable.
5. Any item supplied of inferior quality shall be replaced without any extra charge.
6. The material should be supplied on F.O.R. basis.
7. Failure to supply the items within stipulated time shall be viewed seriously.
8. The payment shall be made after the receipt of material / Instruments.
9. Delivery should be within 30 days and validity of offer will be 60 days.

The Competent Authority, DPCC reserve the all right to reject or accept any quotation without any reason thereof.

N. Moitra
(Dr. Nandita Moitra)
Senior Scientist (Water Lab)

Dr. NANDITA MOITRA
Senior Scientist
Delhi Pollution Control Committee
5th Floor, ISBT Building,
Kashmere Gate, New Delhi-110006